Art Unit: 3724

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Claims 9-16 were withdrawn without traverse in response to restriction requirement on 09/06/05. Therefore, claims 9-16 are cancelled.

The application has been amended as follows:

In the claim:

Claims 9-16 have been cancelled.

Reasons for Allowance

2. The following is an examiner's statement of reasons for allowance: The claims are allowable because prior art fails to teach the step of providing a distance measuring device system for measuring the spacing between the trimmed surface of the specimen and the specimen holder during the movement of the trimming knife o the milling. The prior art also fails to teach the step of transferring the spacing to the microtome or ultramicrotome via a data transfer means prior to mounting the specimen with the specimen holder in the microtome or ultramicrotome as set forth in claim 1.

Regarding claim 1, Gordon (5,533,342) teaches a method of cutting a specimen having a trimmed surface with a cutter. Gordon also teaches that the method include a step of providing a knife holder for the knife and a specimen holder for holding the specimen.

Gordon also teaches a step of using a feed device for producing relative motion between the

Art Unit: 3724

knife and the specimen. Gordon also teaches a step of ascertaining in a trimming apparatus the spacing between the trimmed surface of the specimen and the specimen holder. Gordon also teaches steps of transferring the spacing data to the cutting device and inserting the specimen holder into the cutting device. However, Gordon fails to teach the step of providing a distance measuring device system for measuring the spacing between the trimmed surface of the specimen and the specimen holder during the movement of the trimming knife o the milling. Gordon also fails to teach the step of transferring the spacing to the microtome or ultramicrotome via a data transfer means prior to mounting the specimen with the specimen holder in the microtome or ultramicrotome as set forth in claim 1.

None of these references by themselves or in combination with the other prior art cited teach the claimed invention set forth in claim 1.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Comment

3. In view of amendment to claims 1 and 5, the objection to the drawings under 37 CFR 1.83 (a) and the rejection of claims under 35 U.S.C. 112, second paragraph, have been withdrawn.

Art Unit: 3724

Conclusion

4. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Ghassem Alie whose telephone number is (571) 272-4501. The examiner can normally be reached on Mon-Fri 8:30 am - 5:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Allan Shoap can be reached on (571) 272-4514. The fax phone number for the organization where this application or proceeding is assigned is (571) 273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, SEE http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (too-free).

GA/ga

Allan N. Shoap Supervisory Patent Examiner Group 3700

April 3, 2006